



CERTIFICATE OF ACCREDITATION

ANSI National Accreditation Board

11617 Coldwater Road, Fort Wayne, IN 46845 USA

This is to certify that

Gage-Lab Products, Inc.

6577 S. Cottonwood St.

Murray, UT 84107

has been assessed by ANAB and meets the requirements of international standard

ISO/IEC 17025:2017

while demonstrating technical competence in the field of

CALIBRATION

Refer to the accompanying Scope of Accreditation for information regarding the types of activities to which this accreditation applies

AC-2823

Certificate Number

ANAB Approval

Certificate Valid Through: 11/27/2021
Version No. 001 Issued: 11/27/2019



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

Gage-Lab Products, Inc.

6577 S. Cottonwood St.
Murray, UT 84107
Tim Zimmerman
801-716-2972

CALIBRATION

Valid to: November 27, 2021

Certificate Number: AC-2823

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Video Measuring Systems X & Y Linearity ¹	600 mm x 600 mm	1.8 μ m	Comparison to Glass Scale
Z Linearity ¹	150 mm	1.7 μ m	Comparison to Step Gage

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-2823.

Vice President